

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

NOV 20 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130305</u> ✓
Name <u>Andre Achue</u>		ID# <u>2385</u> Date <u>2-27-09</u> ✓
A	Agency <u>Bethel Police Dept</u>	Phone # <u>907-543-3781</u>
Instrument Location <u>BPD SQUAD ROOM</u> ✓		
Alco S/N <u>58427</u> ✓ Target Value <u>.081</u> High Pressure <u>0.080</u>		
B	Alco Test Values <u>1.084</u> ✓ <u>1.085</u> ✓ 1 st Alco 2 nd Alco	
Signature <u>[Signature]</u> ✓		
(OVER)		

BMB
12/11/09

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 10th day of Dec., 2009.

[Signature]
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

30305

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 170305

FEBRUARY 27, 2009

OPERATOR'S NAME:

ACHEE/ANDRE

OPERATOR'S NUMBER: 2385

SUBJECT'S LAST NAME:

TEST

SUBJECT'S FIRST NAME/MI :

UOC

O.L. #: 1234567

DEPT/AGENCY: BET1

CASE/REPORT: 09-012709

TEST TYPE: T

ALCO TARGET VALUE: .081

ALCO S/N: 58427

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 30.17 in

ALCO TARGET	.081	16:37
BLANK TEST	.000	16:38
INTERNAL STANDARD	VERIFIED	16:38
ALCO TU 30.17 in	.084	16:39
BLANK TEST	.000	16:40
SUBJECT SAMPLE	.080	16:40
BLANK TEST	.000	16:41
ALCO TU 30.17 in	.085	16:41
BLANK TEST	.000	16:42

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 170305

FEBRUARY 27, 2009

TIME 16:32

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 48c

BREATH TUBE: 41c

BAROMETER: 30.17 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@AB CDEF G
HIJKL MNOPQRSTUVWXYZ[\] ^ _ ` abcdefghijklmno
pqrstuvwxyz{|}~